

CERTIFICATE OF CALIBRATION



Deva Medical
1 Chandlers Court, Picow Farm Road, Runcorn, Cheshire
Tel 01928 567571
calibration@deva-medical.co.uk
www.deva-medical.co.uk



9385

Certificate No: UKAS 21063
Date Of Issue: 03 January 2024
Date Of Calibration: 03 January 2024
Issue No: 1
Issued By: N Gargan
Page: 1 of 1

Authorised Signatory

N Gargan

Customer: Deva Calibration Lab
Equipment Set 29

Job Number: 21063

Procedure Number: ICM005 Rev A

Test Conditions:

Ambient Temperature: 22 °C

Calibration Results:

Datalogger	Serial Number	-40°C Range		0°C Range		70°C Range	
		Applied	Indicated	Applied	Indicated	Applied	Indicated
Comark N2011	05040255	-39.95	-40.1	-0.57	-0.5	69.45	68.6
Comark N2011	05040258	-39.95	-40.1	-0.57	-0.6	69.45	69.2
Comark N2011	05040262	-39.95	-39.8	-0.57	-0.6	69.45	66.6
Comark N2011	05040265	-39.95	-39.6	-0.57	-0.6	69.45	67.0
Comark N2011	07040063	-39.95	-39.6	-0.57	-0.4	69.45	68.5
Comark N2011	07040066	-39.95	-39.6	-0.57	-0.6	69.45	67.5
Comark N2011	07040067	-39.95	-39.5	-0.57	-0.6	69.45	67.9
Comark N2011	07040068	-39.95	-39.6	-0.57	-0.4	69.45	68.8
Comark N2011	07040071	-39.95	-39.7	-0.57	-0.6	69.45	67.5
Comark N2011	07040077	-39.95	-39.8	-0.57	-0.6	69.45	68.6
Comark N2014	12190659	-39.95	-39.4	-0.57	-0.5	69.45	68.8

The measurement uncertainty on this certificate is ± 0.24 °C

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service.

It provides traceability of measurement to the SI system of units and / or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%