

CERTIFICATE OF CALIBRATION



Deva Medical
1 Chandlers Court, Picow Farm Road, Runcorn, Cheshire
Tel 01928 567571
calibration@devamedical.co.uk
www.devamedical.co.uk



9385

Certificate No: UKAS 38611
Date Of Issue: 28 October 2025
Date Of Calibration: 28 October 2025
Issue No: 1
Issued By: M Rigby
Page: 1 of 1

Authorised Signatory

N Gargan

Customer: Deva Calibration Lab
Equipment: Set 07
Calibration Location: Deva Calibration Lab
Job Number: 38611

Procedure Number: ICM005 Rev C

Test Conditions:

Ambient Temperature: 22 °C

**The results stated on this certificate relate only to the equipment listed above*

Calibration Results:

Datalogger	Serial Number	-40°C Range		0°C Range		70°C Range	
		Applied	Indicated	Applied	Indicated	Applied	Indicated
Comark N2011	01110080	-39.58	-39.8	0.57	0.5	71.91	72.1
Comark N2011	01110081	-39.58	-39.9	0.57	0.5	71.91	72.0
Comark N2011	01110089	-39.58	-39.9	0.57	0.5	71.91	72.1
Comark N2011	11100184	-39.58	-39.8	0.57	0.6	71.91	72.1
Comark N2011	04210855	-39.58	-39.9	0.57	0.5	71.91	72.0
Comark N2011	04210861	-39.58	-39.8	0.57	0.5	71.91	72.0
Comark N2011	05181030	-39.58	-39.6	0.57	0.6	71.91	72.0
Comark N2011	09120864	-39.58	-39.4	0.57	0.7	71.91	72.0
Comark N2011	12100484	-39.58	-39.6	0.57	0.6	71.91	72.1
Comark N2011	12100504	-39.58	-39.5	0.57	0.6	71.91	72.1
Comark N2014	09230553	-39.58	-39.4	0.57	0.6	71.91	72.1

The measurement uncertainty on this certificate is ± 0.27 °C

The Logger Set is placed in a Climate Chamber and ran over the desired temperature range(s) for approximately 4 hours per temperature range. The data from the logger is downloaded and inserted in the Data Template. The data recorded by the UKAS certified Picolog Thermometer is also downloaded and inserted to the same template. The Data Template will calculate the Mean Value for the Applied and Indicated Temperature, and produce a Certificate.

END OF RESULTS

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service.

It provides traceability of measurement to the SI system of units and / or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%