

# CERTIFICATE OF CALIBRATION



Deva Medical  
1 Chandlers Court, Picow Farm Road, Runcorn, Cheshire  
Tel 01928 567571  
calibration@devamedical.co.uk  
www.devamedical.co.uk



9385

**Certificate No:** UKAS 38614  
**Date Of Issue:** 28 October 2025  
**Date Of Calibration:** 28 October 2025  
**Issue No:** 1  
**Issued By:** M Rigby  
**Page:** 1 of 1

Authorised Signatory

N Gargan

**Customer:** Deva Calibration Lab  
**Equipment:** Set 16  
**Calibration Location:** Deva Calibration Lab  
**Job Number:** 38614

**Procedure Number:** ICM005 Rev C

## Test Conditions:

**Ambient Temperature:** 22 °C

*\*The results stated on this certificate relate only to the equipment listed above*

## Calibration Results:

Datalogger	Serial Number	-40°C Range		0°C Range		70°C Range	
		Applied	Indicated	Applied	Indicated	Applied	Indicated
Comark N2011	05140438	-40.37	-40.7	0.02	0.2	71.70	71.9
Comark N2011	05140619	-40.37	-40.7	0.02	0.1	71.70	71.8
Comark N2011	09230868	-40.37	-40.6	0.02	0.1	71.70	71.7
Comark N2011	07180371	-40.37	-40.5	0.02	0.1	71.70	71.7
Comark N2011	07180372	-40.37	-40.4	0.02	0.1	71.70	71.9
Comark N2011	08110531	-40.37	-40.5	0.02	0.3	71.70	71.9
Comark N2011	09120683	-40.37	-40.3	0.02	0.3	71.70	71.6
Comark N2011	09120685	-40.37	-40.4	0.02	0.1	71.70	71.4
Comark N2011	09120689	-40.37	-40.3	0.02	0.1	71.70	71.3
Comark N2011	09120747	-40.37	-40.4	0.02	0.1	71.70	71.6
Comark N2014	09120947	-40.37	-39.9	0.02	0.2	71.70	71.8

The measurement uncertainty on this certificate is  $\pm 0.27$  °C

The Logger Set is placed in a Climate Chamber and ran over the desired temperature range(s) for approximately 4 hours per temperature range. The data from the logger is downloaded and inserted in the Data Template. The data recorded by the UKAS certified Picolog Thermometer is also downloaded and inserted to the same template. The Data Template will calculate the Mean Value for the Applied and Indicated Temperature, and produce a Certificate.

## END OF RESULTS

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service.*

*It provides traceability of measurement to the SI system of units and / or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.*

*The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%*