

CERTIFICATE OF CALIBRATION



Deva Medical
1 Chandlers Court, Picow Farm Road, Runcorn, Cheshire
Tel 01928 567571
calibration@deva-medical.com
www.deva-medical.com



9385

Certificate No: UKAS 26802
Date Of Issue: 02 August 2024
Date Of Calibration: 02 August 2024
Issue No: 1
Issued By: N GARGAN
Page: 1 of 1

Authorised Signatory

N Gargan

Customer: Deva Calibration Lab
Equipment Set 25

Job Number: 26802

Procedure Number: ICM005 Rev A

Test Conditions:

Ambient Temperature: 22 °C

Calibration Results:

Datalogger	Serial Number	-40°C Range		0°C Range		70°C Range	
		Applied	Indicated	Applied	Indicated	Applied	Indicated
Comark N2011	07180200	-39.59	-39.8	0.56	0.5	71.94	72.1
Comark N2011	07180201	-39.59	-39.7	0.56	0.6	71.94	72.1
Comark N2011	07180203	-39.59	-40.0	0.56	0.5	71.94	72.0
Comark N2011	07180204	-39.59	-39.7	0.56	0.6	71.94	72.0
Comark N2011	07180220	-39.59	-39.8	0.56	0.5	71.94	72.0
Comark N2011	07180262	-39.59	-39.7	0.56	0.6	71.94	72.1
Comark N2011	07180277	-39.59	-40.0	0.56	0.5	71.94	72.1
Comark N2011	07180284	-39.59	-39.8	0.56	0.6	71.94	72.1
Comark N2011	07180299	-39.59	-39.7	0.56	0.5	71.94	72.1
Comark N2011	07180314	-39.59	-39.8	0.56	0.5	71.94	72.1
Comark N2014	03080111	-39.59	-39.0	0.56	0.5	71.94	72.1

The measurement uncertainty on this certificate is ± 0.24 °C

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service.

It provides traceability of measurement to the SI system of units and / or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%