

CERTIFICATE OF CALIBRATION



Deva Medical
1 Chandlers Court, Picow Farm Road, Runcorn, Cheshire
Tel 01928 567571
calibration@deva-medical.com
www.deva-medical.com



9385

Certificate No: UKAS 27472
Date Of Issue: 06 September 2024
Date Of Calibration: 06 September 2024
Issue No: 1
Issued By: N GARGAN
Page: 1 of 1

Authorised Signatory

N Gargan

Customer: Deva Calibration Lab
Equipment Set 13

Job Number: 27472

Procedure Number: ICM005 Rev A

Test Conditions:

Ambient Temperature: 22 °C

Calibration Results:

Datalogger	Serial Number	-40°C Range		0°C Range		70°C Range	
		Applied	Indicated	Applied	Indicated	Applied	Indicated
Comark N2011	05140579	-39.61	-40.0	0.55	0.4	71.88	72.0
Comark N2011	09120715	-39.61	-39.6	0.55	0.6	71.88	71.8
Comark N2011	09120745	-39.61	-39.9	0.55	0.5	71.88	72.0
Comark N2011	09120754	-39.61	-39.6	0.55	0.6	71.88	71.9
Comark N2011	09120755	-39.61	-39.7	0.55	0.6	71.88	72.0
Comark N2011	09120767	-39.61	-39.4	0.55	0.6	71.88	71.9
Comark N2011	09120771	-39.61	-39.7	0.55	0.5	71.88	72.0
Comark N2011	09120773	-39.61	-39.6	0.55	0.6	71.88	72.0
Comark N2011	09120776	-39.61	-39.6	0.55	0.6	71.88	72.0
Comark N2011	09120779	-39.61	-39.7	0.55	0.6	71.88	72.0
Comark N2014	03130362	-39.61	-39.4	0.55	0.8	71.88	71.7

The measurement uncertainty on this certificate is ± 0.24 °C

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service.

It provides traceability of measurement to the SI system of units and / or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%